

CERTIFICATE OF CALIBRATION

ISSUED BY: INSTRON CALIBRATION LABORATORY

DATE OF ISSUE:
See signature

CERTIFICATE NUMBER:
062031425112024



Instron

825 University Avenue
Norwood, MA 02062-2643
Telephone: (800) 473-7838
Fax: (781) 575-5750
Email: service_requests@instron.com

Page 1 of 5 pages

APPROVED SIGNATORY

Type of Calibration: **Displacement**

Relevant Standard: **ASTM E2309/E2309M-20**

Date of Calibration: **14-Mar-2025**

Customer Requested Due Date: **14-Mar-2026**

* * * VERIFICATION RESULTS * * *

System ID: 34SC1B32396

Indicator - Service Port (mm)

Range: 10.0186 mm to 100.1088 mm - Ascending

PASSED Class A

Starting Position: 430.00 mm

Maximum Error: -0.306%

Range: -10.0246 mm to -100.1214 mm - Descending

PASSED Class A

Starting Position: 560.00 mm

Maximum Error: -0.194%

The starting position is measured from the base beam to the bottom of the crosshead.

Customer

Name: Evergreen State College
Address: Science Support Center - NIST
2713 McCann Plaza Drive NW
Olympia, WA 98505
United States
Contact: Carri LeRoy
Email: leroyc@evergreen.edu
Service Order No.: SV2412100254@@@1

Machine/System

Manufacturer: Instron
Serial No.: 34SC1B32396
Condition: Good

Temperature

Starting Temperature: 67.2 °F
Final Temperature: 67.4 °F

Methodology

The assessment of the testing machine was conducted on site at the above customer location in accordance with ASTM E2309/E2309M-20 "Standard Practices for Verification of Displacement Measuring Systems and Devices Used in Material Testing Machines" (Follow-the-Displacement Method) using Instron procedure ICA-8-07.

The system was calibrated in the 'As Found' condition with no adjustments or repairs carried out. This is also the 'As Left' condition.

Prior to verification, a pre-calibration inspection was conducted. During the inspection, the testing system was found to be in Good condition.

The calibrated range of displacement includes only those displacements which are greater than or equal to the ASTM Lower Limit.

Three calibration tests were made with the testing machine in the vertical position.

System Classification

The calibration and equipment used conform to a controlled Quality Assurance program which meets the specifications

Instron CalproCR Version 3.58

The results indicated on this certificate and the following report relate only to the items calibrated. If there are methods or data included that are not covered by the NVLAP accreditation it will be identified in the comments. Any limitations of use as a result of this calibration will be indicated in the comments. This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the U.S. Government. This report shall not be reproduced, except in full, without the approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

NVLAP ACCREDITED CALIBRATION LABORATORY No. 200301-0

CERTIFICATE NUMBER:

062031425112024

Page 2 of 5 pages

outlined in ANSI/NCSL Z540.1-1994, ISO 10012:2003, ISO 9001:2015, ISO/IEC 17025:2017.

The displacement-measuring system has been verified for the displacements indicated using equipment calibrated within the requirements of ASTM E2309/E2309M-20.

The Simple Acceptance decision rule has been agreed to and employed in the determination of conformance to the identified metrological specification.

Classifications based upon ASTM E2309 Table 1.

| Classification | Resolution not to exceed the greater of: | | Error/Repeatability not to exceed the greater of: | |
|----------------|--|--------------|---|---------------------------------------|
| | Fixed Error mm [in] | % of Reading | Fixed Error mm [in] | Relative Error (% of Displacement) |
| Class A | 0.013 [0.0005] | +/- 0.25 | +/- 0.025 [0.001] | +/- 0.5 |
| Class B | 0.038 [0.0015] | +/- 0.5 | +/- 0.075 [0.003] | +/- 1.0 |
| Class C | 0.064 [0.0025] | +/- 1.0 | +/- 0.125 [0.005] | +/- 2.0 |
| Class D | 0.13 [0.005] | +/- 1.5 | +/- 0.25 [0.010] | +/- 3.0 |

Note: Classification is based upon meeting both Resolution and Error/Repeatability requirements.

Data Summary - Indicator - Service Port (mm)

| % of Range | Run 1 Error | | | Run 2 Error | | | Run 3 Error | | | Repeat Error | |
|---|-------------|--------|-------|-------------|--------|-------|-------------|--------|-------|--------------|-------|
| | (mm) | (%) | Class | (mm) | (%) | Class | (mm) | (%) | Class | (mm) | Class |
| Range: 10.0186 mm to 100.1088 mm - Ascending | | | | | | | | | | | |
| 10 | -0.0165 | -0.164 | A | -0.0307 | -0.306 | A | -0.0306 | -0.305 | A | 0.0142 | A |
| 25 | -0.0284 | -0.113 | A | -0.0455 | -0.182 | A | -0.0505 | -0.202 | A | 0.0221 | A |
| 50 | -0.0471 | -0.094 | A | -0.0481 | -0.096 | A | -0.0673 | -0.134 | A | 0.0202 | A |
| 75 | -0.0652 | -0.087 | A | -0.0826 | -0.110 | A | -0.0871 | -0.116 | A | 0.0219 | A |
| 100 | -0.0782 | -0.078 | A | -0.0979 | -0.098 | A | -0.0998 | -0.100 | A | 0.0216 | A |
| Range: -10.0246 mm to -100.1214 mm - Descending | | | | | | | | | | | |
| 10 | 0.0179 | -0.178 | A | 0.0184 | -0.184 | A | 0.0194 | -0.194 | A | 0.0015 | A |
| 25 | 0.0390 | -0.156 | A | 0.0390 | -0.156 | A | 0.0418 | -0.167 | A | 0.0028 | A |
| 50 | 0.0661 | -0.132 | A | 0.0642 | -0.128 | A | 0.0634 | -0.127 | A | 0.0027 | A |
| 75 | 0.0901 | -0.120 | A | 0.0835 | -0.111 | A | 0.0823 | -0.110 | A | 0.0078 | A |
| 100 | 0.1125 | -0.112 | A | 0.1059 | -0.106 | A | 0.1052 | -0.105 | A | 0.0073 | A |

Data - Indicator - Service Port (mm)

| % of Range | Run 1 | | Run 2 | | Run 3 | | Uncertainty of Measurement* | |
|--|-------------------|-----------------|-------------------|-----------------|-------------------|-----------------|--------------------------------|----------|
| | Indicated (mm) | Applied (mm) | Indicated (mm) | Applied (mm) | Indicated (mm) | Applied (mm) | % | (+/- mm) |
| Range: 10.0186 mm to 100.1088 mm - Ascending | | | | | | | | |
| 0 | 0.00000 | 0.0000 | 0.00000 | 0.0000 | 0.00013 | 0.0000 | | |
| 10 | 10.00213 | 10.0186 | 10.00750 | 10.0382 | 10.00713 | 10.0376 | 0.066 | 0.0067 |
| 25 | 25.00600 | 25.0344 | 25.00850 | 25.0540 | 25.00663 | 25.0570 | 0.039 | 0.0097 |
| 50 | 50.00888 | 50.0560 | 50.02488 | 50.0730 | 50.00462 | 50.0718 | 0.033 | 0.016 |
| 75 | 75.00938 | 75.0746 | 75.00762 | 75.0902 | 75.00525 | 75.0922 | 0.031 | 0.024 |
| 100 | 100.00525 | 100.0834 | 100.00850 | 100.1064 | 100.00913 | 100.1088 | 0.031 | 0.031 |

CERTIFICATE OF CALIBRATION

NVLAP ACCREDITED CALIBRATION LABORATORY No. 200301-0

CERTIFICATE NUMBER:

062031425112024

Page 3 of 5 pages


Data - Indicator - Service Port (mm)

| % of Range | Run 1 | | Run 2 | | Run 3 | | Uncertainty of Measurement* | |
|---|----------------|--------------|----------------|--------------|----------------|--------------|-----------------------------|----------|
| | Indicated (mm) | Applied (mm) | Indicated (mm) | Applied (mm) | Indicated (mm) | Applied (mm) | % | (+/- mm) |
| Range: -10.0246 mm to -100.1214 mm - Descending | | | | | | | | |
| 0 | 0.00000 | 0.0000 | -0.00025 | 0.0000 | 0.00000 | 0.0000 | | |
| 10 | -10.00675 | -10.0246 | -10.00825 | -10.0264 | -10.00838 | -10.0278 | 0.066 | 0.0067 |
| 25 | -25.00825 | -25.0472 | -25.00588 | -25.0446 | -25.00463 | -25.0464 | 0.039 | 0.0096 |
| 50 | -50.00512 | -50.0712 | -50.00825 | -50.0722 | -50.00600 | -50.0694 | 0.033 | 0.016 |
| 75 | -75.00650 | -75.0966 | -75.00574 | -75.0890 | -75.00713 | -75.0894 | 0.031 | 0.024 |
| 100 | -100.00887 | -100.1214 | -100.00775 | -100.1134 | -100.00500 | -100.1102 | 0.031 | 0.031 |

* The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor, $k = 2$, providing a level of confidence of approximately 95%.

The uncertainty stated refers to values obtained during the calibration and makes no allowances for factors such as long-term drift, temperature and alignment effects - the influence of such factors should be taken into account.

Graphical Data - Indicator - Service Port (mm)



CERTIFICATE OF CALIBRATION


NVLAP ACCREDITED CALIBRATION LABORATORY No. 200301-0

CERTIFICATE NUMBER:

062031425112024

Page 4 of 5 pages

Graphical Data - Indicator - Service Port (mm)



Calibration Equipment

The measurement results produced with Instron standards are traceable to the SI (The International System of Units) through internationally recognized National Metrology Institutes (NIST, NPL, PTB, Inmetro, etc.).

| Manufacturer/Model | Serial No. | Description | Cal Date | Cal Due | Certificate Ref. |
|-----------------------|----------------|-----------------|-------------|-------------|------------------|
| Instron LDS (280/287) | 051613C (ASTM) | disp. indicator | 04-Jun-2024 | 04-Jun-2026 | INS051613C |
| Extech 445580 | 1103444 | temp. indicator | 25-Jul-2024 | 25-Jul-2025 | 2024016106 |

The class of the calibration equipment was equal to or better than the class to which this testing machine has been calibrated.

Calibration Equipment Usage

| Measurement Type | Serial No. | Direction | Percent(s) of Range | Accuracy (+/-) |
|------------------|----------------|------------|---------------------|----------------|
| Displacement | 051613C (ASTM) | Ascending | 10/ 25/ 50/ 75/ 100 | 0.0002 in |
| | 051613C (ASTM) | Descending | 10/ 25/ 50/ 75/ 100 | 0.0002 in |
| Temperature | 1103444 | All | All | 1.8 °F |

The accuracy of the calibration equipment used was equal to or better than the accuracy indicated in the table above.

Comments

New Installation

CERTIFICATE OF CALIBRATION

NVLAP ACCREDITED CALIBRATION LABORATORY No. 200301-0

CERTIFICATE NUMBER:

062031425112024

Page 5 of 5 pages

Performed By: Mike Spuzzillo
Field Service Engineer

It is Instron's recommendation that the displacement measuring system be calibrated at least annually or after any repair or adjustment which affects the accuracy of measurements.